

ENG

Product Data Sheet:

**HUMAN CALCITONIN (CT)
ELISA, HIGH SENSITIVITY**

Catalogue number:

RIS0019R

For research use only!

BioVendor – Laboratorní medicína a.s.

Karásek 1767/1, 621 00 Brno, Czech Republic

+420 549 124 185

info@biovendor.com

sales@biovendor.com

www.biovendor.com

1. INTENDED USE	3
2. STORAGE, EXPIRATION	3
3. INTRODUCTION	3
4. TEST PRINCIPLE	4
5. PRECAUTIONS	4
6. TECHNICAL HINTS	5
7. REAGENT SUPPLIED	5
8. MATERIAL REQUIRED BUT NOT SUPPLIED	7
9. PREPARATION OF REAGENTS	7
10. PREPARATION OF SAMPLES	8
11. ASSAY PROCEDURE	8
12. CALCULATIONS	9
13. PERFORMANCE CHARACTERISTICS	10
14. QUALITY CONTROL	12
15. REFERENCE INTERVALS	12
16. REFERENCES	13
17. ASSAY PROCEDURE SUMMARY	14

1. INTENDED USE

The RIS0019R Human Calcitonin (CT) ELISA, High sensitivity Immunoenzymetric assay for the measurement of human Calcitonin in serum. It is intended for research use only.

2. STORAGE, EXPIRATION

Before opening or reconstitution, all kits components are stable until the expiry date, indicated on the vial label, if kept at 2 to 8°C.

Unused wells must be stored, at 2-8°C, in a sealed bag containing a desiccant until expiration date.

After reconstitution, calibrators, controls and CT free serum should be frozen immediately after use and kept at -20°C for 3 months. Only one freeze thawing cycle is allowed, discard the calibrators, controls and CT free serum after the second use.

The concentrated Wash Solution is stable at room temperature until expiration date.

Freshly prepared Working Wash solution should be used on the same day.

After its first use, the concentrated conjugate (50x) is stable until expiry date, if kept in the original well-closed vial at 2 to 8°C.

The Working anti-CT-HRP conjugate is stable for 1 week at 4°C.

The chromogenic TMB solution and the Stop Solution are stable at 2°C to 8°C until the expiry date.

Alterations in physical appearance of kit reagents may indicate instability or deterioration.

3. INTRODUCTION

Biological activities

Calcitonin (CT) is a 32 amino acid peptide hormone secreted by the para-follicular C-cells of the thyroid gland under serum calcium control. After acute administration, this peptide acts as a potent hypocalcemic and hypophosphatemic hormone by increasing renal calcium clearance and reducing bone resorption. However, its precise physiological role in bone metabolism is not yet fully understood.

Various forms of CT may be detected in blood samples, including a CT monomer, an oxidized monomer, a dimer, higher molecular weight forms, and possibly precursor of CT. The concentrations of these peptides vary with clinical status, renal function and tissular origin of CT (normal or ectopic production).

Medullar thyroid carcinoma (MTC) is a malignant tumor, developed from the C-cells, secreting calcitonin in large excess. This disease occurs either as a sporadic (80%) or a familial (20%) form, which is transmitted as an autosomal dominant gene or as a component of multiple endocrine neoplasia (IIb).

Moderate hypercalcitoninemia is also observed in pregnancy, pernicious anaemia, renal failure and during the neonatal period. Preferably, monomer form of CT is detected in this assay.

Clinical Application

The measurement of CT is used for:

- Diagnosis of medullary thyroid carcinoma (MTC),
- Follow up of malignant tumors, to check the success of surgery and to monitor for recurrence,
- Diagnosis of the preclinical cases of the familial forms of MTC (MEN II or Sipple syndrome) by the use of stimulation tests (calcium or pentagastrin),
- Study of the pathophysiology of the calcium-phosphate and bone metabolism.

4. TEST PRINCIPLE

The Biovendor Calcitonin ELISA is a solid phase Enzyme Amplified Sensitivity Immunoassay performed on microtiterplate. Calibrators and samples react with the capture monoclonal antibody (MAb 1) coated on microtiter well and with a monoclonal antibody (MAb 2) labelled with horseradish peroxidase (HRP). After an incubation period allowing the formation of a sandwich: coated MAb 1 – human CT – MAb 2 – HRP, the microtiterplate is washed to remove unbound enzyme labelled antibody. Bound enzyme-labelled antibody is measured through a chromogenic reaction. Chromogenic solution (TMB ready for use) is added and incubated. The reaction is stopped with the addition of Stop Solution and the microtiterplate is then read at the appropriate wavelength. The amount of substrate turnover is determined colourimetrically by measuring the absorbance, which is proportional to the Calcitonin concentration. A calibration curve is plotted and Calcitonin concentration in samples is determined by interpolation from the calibration curve.

5. PRECAUTIONS

For research use only

The human blood components included in this kit have been tested by European approved and/or FDA approved methods and found negative for HBsAg, anti-HCV, anti-HIV-1 and 2. No known method can offer complete assurance that human blood derivatives will not transmit hepatitis, AIDS or other infections. Therefore, handling of reagents, serum specimens should be in accordance with local safety procedures.

All animal products and derivatives have been collected from healthy animals. Bovine components originate from countries where BSE has not been reported. Nevertheless, components containing animal substances should be treated as potentially infectious.

Avoid any skin contact with all reagents, Stop Solution contains HCl, the chromogenic solution contains TMB and H₂O₂. In case of contact, wash thoroughly with water.

Do not smoke, drink, eat or apply cosmetics in the working area. Do not pipette by mouth. Use protective clothing and disposable gloves.

6. TECHNICAL HINTS

Do not use the kit or components beyond expiry date.

Do not mix materials from different kit lots.

Bring all the reagents to room temperature prior to use.

Thoroughly mix all reagents and samples by gentle agitation or swirling.

Perform calibrators, controls and samples in duplicate. Vertical alignment is recommended.

Use a clean plastic container to prepare the Wash Solution.

In order to avoid cross-contamination, use a clean disposable pipette tip for the addition of each reagent and sample.

For the dispensing of the Chromogenic Solution and the Stop Solution avoid pipettes with metal parts.

High precision pipettes or automated pipetting equipment will improve the precision.

Respect the incubation times.


Prepare a calibration curve for each run, do not use data from previous runs.

The chromogenic solution should be colourless. If a dark blue colour develops within a few minutes after preparation, this indicates that the preparation is unusable and must be discarded.

Dispense the Chromogenic Solution within 15 minutes following the washing of the microtiterplate.

During incubation with Chromogenic Solution, avoid direct sunlight on the microtiterplate.

7. REAGENT SUPPLIED

Reagents	96 tests Kit	Color Code	Reconstitution			
 Microtiterplate with 96 anti CT (monoclonal antibodies) coated breakable wells	96 wells	blue	Ready for use			
<table border="1" data-bbox="177 1700 505 1778"> <tr> <td>Ab</td> <td>HRP</td> <td>CONC</td> </tr> </table> Conjugate: HRP labelled anti-CT (monoclonal antibodies) in Stabilizing Buffer	Ab	HRP	CONC	1 vial 0.125 ml	yellow	Dilute 50 x with conjugate buffer
Ab	HRP	CONC				

<table border="1" data-bbox="177 192 384 264"> <tr> <td>CONJ</td> <td>BUF</td> </tr> </table> <p>Conjugate buffer: TRIS-Maleate buffer with bovine serum albumin, EDTA and thymol</p>	CONJ	BUF	1 vial 6 ml	red	Ready for use	
CONJ	BUF					
<table border="1" data-bbox="161 421 341 483"> <tr> <td>CAL</td> <td>N</td> </tr> </table> <p>Calibrator N = 0 to 5 (see exact values on vial labels) in CT-Free human serum</p>	CAL	N	6 vials lyophilized	yellow	Add 0.5 ml distilled water	
CAL	N					
<table border="1" data-bbox="177 696 504 763"> <tr> <td>CONTROL</td> <td>N</td> </tr> </table> <p>Controls - N = 1 or 2 in human serum with gentamycin</p>	CONTROL	N	2 vials lyophil.	silver	Add 0.5 ml distilled water	
CONTROL	N					
<table border="1" data-bbox="161 920 296 981"> <tr> <td>SERUM</td> </tr> </table> <p>CT free human serum (to be used for samples dilution) with thymol</p>	SERUM	1 vial lyophil.	black	Add buffer (see reconstitution volume in the Quality Control Sheet)		
SERUM						
<table border="1" data-bbox="161 1133 296 1193"> <tr> <td>BUF</td> </tr> </table> <p>Buffer (serum free): borate buffer</p>	BUF	1 vial 8 ml	black	Ready for use		
BUF						
<table border="1" data-bbox="217 1312 563 1373"> <tr> <td>WASH</td> <td>SOLN</td> <td>CONC</td> </tr> </table> <p>Wash Solution (Tris-HCl)</p>	WASH	SOLN	CONC	1 vial 10 ml	brown	Dilute 200 x with distilled water (use a magnetic stirrer).
WASH	SOLN	CONC				
<table border="1" data-bbox="217 1473 504 1563"> <tr> <td>CHROM</td> <td>TMB</td> </tr> </table> <p>Chromogenic TMB solution (Tetramethylbenzidine)</p>	CHROM	TMB	1 vial 12 ml	brown	Ready for use	
CHROM	TMB					
<table border="1" data-bbox="217 1693 424 1753"> <tr> <td>STOP</td> <td>SOLN</td> </tr> </table> <p>Stop solution: HCl 1.0 N</p>	STOP	SOLN	1 vial 12 ml	white	Ready for use	
STOP	SOLN					

- Note:**
1. CT free human serum is to be used for samples dilution.
 2. 1 pg of our reference preparation is equivalent to 0.19 μ IU NIBSC 89/620.

8. MATERIAL REQUIRED BUT NOT SUPPLIED

1. High quality distilled water
2. Pipettes for delivery of: 50 μ l, 100 μ l, 500 μ l and 1 ml (the use of accurate pipettes with disposable plastic tips is recommended)
3. Vortex mixer
4. Magnetic stirrer
5. Washer for microtiterplate
6. Microtiterplate reader capable of reading at 450 nm and 650 nm (bichromatic reading).

9. PREPARATION OF REAGENTS

Calibrators

Reconstitute the calibrators with 0.5 ml distilled water.

Controls

Reconstitute the controls with 0.5 ml distilled water.

Working Wash solution

Prepare an adequate volume of Working Wash solution by adding 199 volumes of distilled water to 1 volume of Wash Solution (200x). Use a magnetic stirrer to homogenize. Discard unused Working Wash solution at the end of the day.

CT Free Serum

Reconstitute the CT Free Serum with the amount of Buffer as indicated on the Quality Control Sheet. Allow it to remain undisturbed until completely dissolved, and then mix well by gentle inversion.

Working anti-CT-HRP conjugate

Prepare an adequate volume of conjugate solution by adding for example :40 μ l of the 50 x concentrated anti-CT-HRP conjugate to 2 ml of conjugate buffer. Use a vortex to homogenize. Extemporaneous preparation is recommended.

10. PREPARATION OF SAMPLES

Serum must be kept at 2 - 8°C.

If the test is not run within 24 hours, storage in aliquots at -20°C is recommended. Avoid subsequent freeze thaw cycles.

Prior to use, all samples should be at room temperature. It is recommended to vortex the samples before use.

Do not use haemolysed samples.

Do not use lipemic samples.

11. ASSAY PROCEDURE

1. Select the required number of wells for the run. The unused wells should be resealed in the bag with a desiccant and stored at 2-8°C.
2. Secure the wells into the holding frame.
3. Pipette 100 µl of each Calibrator, Control and Sample into the appropriate wells.
4. Pipette 50 µl of Working anti-CT-HRP conjugate into all the wells.
5. Incubate for 18 ± 1 hour at 2-8°C.
6. Aspirate the liquid from each well.
7. Wash the plate 3 times by:
 - Dispensing 0.4 ml of Wash Solution into each well
 - Aspirating the content of each well
8. Pipette 100 µl of the Chromogenic solution into each well within 15 minutes following the washing step.
9. Incubate the microtiterplate for 30 minutes at room temperature avoiding direct sunlight.
10. Pipette 100 µl of Stop Solution into each well.

Read the absorbances at 450 nm (reference filter 630 nm or 650 nm) within 1 hour and calculate the results as described in section 12.

12. CALCULATIONS

1. Read the plate at 450 nm against a reference filter set at 650 nm (or 630 nm).
2. Calculate the mean of duplicate determinations.
3. Plot the OD values (ordinate) for each calibrator against the corresponding concentration of Calcitonin (abscissa) and draw a calibration curve through the calibrator points.
4. Read the concentration for each control and sample by interpolation on the calibration curve.

Computer assisted data reduction will simplify these calculations. If automatic result processing is used, a 4-parameter logistic function curve fitting is recommended.

The following data are for illustration only and should never be used instead of the real time calibration curve.

CT-U.S.-ELISA		Bichromatic model (OD)
Calibrator	0 pg/ml	0.009
	10 pg/ml	0.029
	50 pg/ml	0.127
	100 pg/ml	0.447
	200 pg/ml	0.919
	400 pg/ml	1.87

13. PERFORMANCE CHARACTERISTICS

Limit

Twenty zero calibrators were assayed along with a set of other calibrators. The detection limit, defined as the apparent concentration two standard deviations above the average OD at zero binding, was 0.7 pg/ml.

Specificity

Some potentially interfering hormones have been tested in this assay. At concentrations up to 100 ng/ml, none of the following hormones showed significant interference :

- CGRP
- Salmon-calcitonin
- PDN 21
- Procalcitonin N terminal.

Precision

Intra assay (n=19)

Serum	<X> ± SD (pg/ml)	CV %
A	43.0 ± 0.75	1.7
B	133.7 ± 5.2	3.9

SD : Standard Deviation; CV: Coefficient of variation

Inter assay (n=8)

Serum	<X> ± SD (pg/ml)	CV %
A	44.6 ± 2.1	4.9
B	136.3 ± 8.1	6

SD : Standard Deviation; CV: Coefficient of variation

Accuracy

Recovery test

Added CT (pg/ml)	Recovered CT (pg/ml)	Recovery (%)
327.7	340.6	104
160.7	159.3	99
80.5	80.4	99
48.3	50.8	105

Dilution test

Sample	Dilution	Theoretical Concent. (pg/ml)	Measured Concent. (pg/ml)
Serum 1	1/1	-	300.6
	1/2	150.3	157.9
	1/4	75.1	75.5
	1/8	37.6	45.7
	1/16	18.8	25.2
	1/32	9.4	12.1
	1/64	4.7	5

Samples were diluted with CT Free Serum.

Hook effect

A sample spiked with CT up to 480000 pg/ml gives higher OD's than the last calibrator point.

14. QUALITY CONTROL

- If the results obtained for Control 1 and/or Control 2 are not within the range specified on the vial label, the results cannot be used unless a satisfactory explanation for the discrepancy has been given.
- If desirable, each laboratory can make its own pools of control samples, which should be kept frozen in aliquots. Controls that contain azide will interfere with the enzymatic reaction and cannot be used.
- Acceptance criteria for the difference between the duplicate results of the samples should rely on Good Laboratory Practises
- It is recommended that Controls be routinely assayed as unknown samples to measure assay variability. The performance of the assay should be monitored with quality control charts of the controls.
- It is good practise to check visually the curve fit selected by the computer.

15. REFERENCE INTERVALS

These values are given only for guidance; each laboratory should establish its own normal range of values.

Normal values

84 samples from normal subjects obtained values below 11 pg/ml.

16. REFERENCES

References to calcitonin:

1. GRAZE K., SPILER I.J., TASHIJAN A.H., MELVIN K.E.W., CERVI-SKINNER S., GAGEL R.F., MILLER H.H., WOLFE H.J., DELELLIS R.A., LEAPE L., FELDMAN Z.T. and REICHLIN S. (1978) Natural history of familial medullary thyroid carcinoma; Effect of a program for early diagnosis. *Engl. J. Med.*, 299,18;980-985.
2. HENNESSY J.F., WELLS S.A., ONTJES D.A. and COOPER C.W. (1974) A comparison of pentagastrin injection and calcium infusion as provocative agents for the detection of medullary carcinoma of the thyroid. *J. Clin. Endocrinol. Metab.*, 39:487-495.
3. ROUGIER Ph., CALMETTES C., LAPLANCHE A., TRAVAGLI J.P., LEFEVRE M., PARMENTIER C., MILHAUD G. and TUBIANA M. (1983) the values of calcitonin and carcinoembryonic antigen in the treatment and management of nonfamilial medullary thyroid carcinoma. *Cancer*, 51,5:856-862.
4. WALLACH S.R., ROYSTON I., TAETLE R., WOHL H. and DEFTOS L. (1981) Plasma calcitonin as a marker of disease activity in patients with small cell carcinoma of the lung. *J. Clin. Endocrinol. Metab.*, 53,3:602-606.
5. WELLS S.A., BAYLIN S.B., LINEHAN W.M., FARRELL R.E., COX E.B. and COOPER C.W. (1978) Provocative agents and the diagnosis of medullary carcinoma of the thyroid gland. *Ann. Surg.*, 188,2:139-141.
6. AURBACH G.D., MARX S.J. and SPIEGEL A.M. (1985) Parathyroid hormone, calcitonin, and the calciferols. In: *Williams Textbook of endocrinology* (7th edition; Wilson J.D. and Foster D.W. eds) W.B. Saunders Company, Philadelphia, 1137-1217.
7. BODY J.J. et al. (1987) SCC antigen and other tumor markers in lung cancer: preliminary results. *Excerpta Medica*, 162-170.
8. NICOLI P. et al. (1995) Abnormal calcitonin basal levels and pentagastrin response in patients with chronic renal failure on maintenance hemodialysis. *Eur. J. Endocrinol.* 132, 1, 75-81.
9. PACINI F. et al. (1994) Routine measurement of serum calcitonin in nodular disease allows the preoperative diagnosis of unsuspected sporadic medullary thyroid carcinoma. *J. Clin. Endocrinol. Metab. Excerpta Medica*, 78, 4, 824-9.
10. QUESADA J. M. et al. (1994) Calcitriol corrects deficient calcitonin secretion in the Vit. D deficient elderly. *J. Bone Miner Res.* 9, 1, 53-57.

For more references on this product see our web pages at www.biovendor.com.

17. ASSAY PROCEDURE SUMMARY

	CALIBRATORS (μl)	SAMPLE(S) CONTROLS (μl)
Calibrators (0-5)	100	-
Controls, Samples	-	100
Working Anti-CT-HRP conjugate	50	50
Incubate for 18 ± 1 hours at $2 - 8^{\circ}\text{C}$. Aspirate the contents of each well. Wash 3 times with 400 μ l of Wash Solution and aspirate.		
Chromogenic TMB Solution	100	100
Incubate for 30 min at room temperature.		
Stop Solution	100	100
Read on a microtiterplate reader and record the absorbance of each well at 450 nm (versus 630 or 650 nm)		



BioVendor – Laboratorní medicína a.s.
Karásek 1767/1, 621 00 Brno, Czech Republic
+420 549 124 185
info@biovendor.com
sales@biovendor.com
www.biovendor.com

Example Version

