


BioVendor

Research
and Diagnostic Products



Mouse Interleukin-12 p70 ELISA

Product Data Sheet

Cat. No.: RAF031R

For Research Use Only

CONTENTS

1.	INTENDED USE	3
2.	SUMMARY	3
3.	PRINCIPLES OF THE TEST	4
4.	REAGENTS PROVIDED	5
5.	STORAGE INSTRUCTIONS – ELISA KIT	5
6.	SAMPLE COLLECTION AND STORAGE INSTRUCTIONS	6
7.	MATERIALS REQUIRED BUT NOT PROVIDED	6
8.	PRECAUTIONS FOR USE	7
9.	PREPARATION OF REAGENTS	8
10.	TEST PROTOCOL	10
11.	CALCULATION OF RESULTS	12
12.	LIMITATIONS	15
13.	PERFORMANCE CHARACTERISTICS	15
14.	REAGENT PREPARATION SUMMARY	20
15.	TEST PROTOCOL SUMMARY	21

**➤➤ This kit is manufactured by:
BioVendor – Laboratorní medicína a.s.**

➤➤ Use only the current version of Product Data Sheet enclosed with the kit!

1. INTENDED USE

The Mouse IL-12 p70 ELISA Kit is an enzyme-linked immunosorbent assay for the quantitative detection of mouse IL-12p70.

2. SUMMARY

Interleukin-12 (IL-12) is a pleiotropic cytokine, formerly termed cytotoxic lymphocyte maturation factor (CLMF) or natural killer cell stimulatory factor (NKSF), which is produced primarily by stimulated macrophages. IL-12 has been shown to be a proinflammatory cytokine produced by phagocytic cells, B-cells, and other antigen – presenting cells that modulate adaptive immune responses by favoring the generation of T-helper type 1 cells.

IL-12 exerts a variety of biological effects on T- and natural killer cells. Apart from promotion of Th1 development and its ability to promote cytolytic activity it mediates some of its physiological activities by acting as a potent inducer of interferon (IFN) γ production and the stimulation of other cytokines from peripheral blood T- and NK-cells. IFN γ then enhances the ability of the phagocytic cells to produce IL-12 and other proinflammatory cytokines. Thus, IL-12 induced IFN γ acts in a positive feedback loop that represents an important amplifying mechanism in the inflammatory response to infections.

Its role in directing development of a Th1 type immune response from naive T-cells demonstrates its critical role in regulation of the immune response and strongly suggests its potential usefulness in cancer therapy.


IL-12 is a disulfide-linked heterodimeric cytokine composed of a 35kDa light chain (p35) and a 40kDa heavy chain (p40) resulting in the only biologically active 70kDa (p70) form of IL-12. The p40 subunit can also form a homodimer which has been shown to be able to bind the IL-12 receptor and thus acts as an IL-12 antagonist. Additionally, the p40 subunit has been found to be expressed in a high excess over p70.

The critical role of IL-12 in several pathogeneses has been shown.

3. PRINCIPLES OF THE TEST


An anti-mouse IL-12p70 coating antibody is adsorbed onto microwells.

Figure 1




Mouse IL-12p70 present in the sample or standard binds to antibodies adsorbed to the microwells. A biotin-conjugated anti-mouse IL-12p70 antibody is added and binds to mouse IL-12p70 captured by the first antibody.

Figure 2




Following incubation unbound biotin-conjugated anti-mouse IL-12p70 antibody is removed during a wash step. Streptavidin-HRP is added and binds to the biotin-conjugated anti-mouse IL-12p70 antibody.

Figure 3




Following incubation unbound Streptavidin-HRP is removed during a wash step, and substrate solution reactive with HRP is added to the wells.

Figure 4



A colored product is formed in proportion to the amount of mouse IL-12p70 present in the sample or standard. The reaction is terminated by addition of acid and absorbance is measured at 450 nm. A standard curve is prepared from 7 mouse IL-12p70 standard dilutions and mouse IL-12p70 sample concentration determined.

Figure 5



4. REAGENTS PROVIDED

- 1 aluminium pouch with an **Antibody Coated Microtiter Strips** (12 strips with 8 wells each) coated with monoclonal antibody to mouse IL-12p70
- 1 vial (70 µl) **Biotin-Conjugate** anti-mouse IL-12 p70 monoclonal antibody
- 1 vial (150 µl) **Streptavidin-HRP**
- 2 vials mouse IL-12p70 **Standard** lyophilized, 2 ng/ml upon reconstitution
- 1 vial (12 ml) **Sample Diluent**
- 1 vial (5 ml) **Assay Buffer Concentrate** 20x (PBS with 1% Tween 20 and 10% BSA)
- 1 bottle (50 ml) **Wash Buffer Concentrate** 20x (PBS with 1% Tween 20)
- 1 vial (15 ml) **Substrate Solution** (tetramethyl-benzidine)
- 1 vial (15 ml) **Stop Solution** (1M Phosphoric acid)
- 4 **Adhesive Films**

5. STORAGE INSTRUCTIONS – ELISA KIT

Store kit reagents between 2°C and 8°C. Immediately after use remaining reagents should be returned to cold storage (2°C to 8°C).

Expiry of the kit and reagents is stated on labels.

Expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, this reagent is not contaminated by the first handling.

6. SAMPLE COLLECTION AND STORAGE INSTRUCTIONS

Cell culture supernatant and serum were tested with this assay. Other biological samples might be suitable for use in the assay. Remove serum from the clot as soon as possible after clotting. Samples containing a visible precipitate must be clarified prior to use in the assay. Do not use grossly hemolyzed or lipemic samples.

Samples should be aliquoted and must be stored frozen at -20°C to avoid loss of bioactive mouse IL-12p70. If samples are to be run within 24 hours, they may be stored at 2°C to 8°C (refer to "Sample stability"). Avoid repeated freeze-thaw cycles. Prior to assay, the frozen sample should be brought to room temperature slowly and mixed gently.

7. MATERIALS REQUIRED BUT NOT PROVIDED

- 5 ml and 10 ml graduated pipettes
- 5 μl to 1000 μl adjustable single channel micropipettes with disposable tips
- 50 μl to 300 μl adjustable multichannel micropipette with disposable tips
- Multichannel micropipette reservoir
- Beakers, flasks, cylinders necessary for preparation of reagents
- Device for delivery of wash solution (multichannel wash bottle or automatic wash system)
- Microwell strip reader capable of reading at 450 nm (620 nm as optional reference wave length)
- Glass-distilled or deionized water
- Statistical calculator with program to perform regression analysis

8. PRECAUTIONS FOR USE

- All reagents should be considered as potentially hazardous. We therefore recommend that this product is handled only by those persons who have been trained in laboratory techniques and that it is used in accordance with the principles of good laboratory practice. Wear suitable protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken to avoid contact with skin or eyes. In the case of contact with skin or eyes wash immediately with water. See material safety data sheet(s) and/or safety statement(s) for specific advice.
- Reagents are intended for research use only and are not for use in diagnostic or therapeutic procedures.
- Do not mix or substitute reagents with those from other lots or other sources.
- Do not use kit reagents beyond expiration date on label.
- Do not expose kit reagents to strong light during storage or incubation.
- Do not pipette by mouth.
- Do not eat or smoke in areas where kit reagents or samples are handled.
- Avoid contact of skin or mucous membranes with kit reagents or samples.
- Rubber or disposable latex gloves should be worn while handling kit reagents or samples.
- Avoid contact of substrate solution with oxidizing agents and metal.
- Avoid splashing or generation of aerosols.
- To avoid microbial contamination or cross-contamination of reagents or samples which may invalidate the test use disposable pipette tips and/or pipettes.
- Use clean, dedicated reagent trays for dispensing the conjugate and substrate reagent.
- Exposure to acid inactivates the conjugate.
- Glass-distilled water or deionized water must be used for reagent preparation.
- Substrate solution must be at room temperature prior to use.
- Decontaminate and dispose samples and all potentially contaminated materials as they could contain infectious agents. The preferred method of decontamination is autoclaving for a minimum of 1 hour at 121.5°C.
- Liquid wastes not containing acid and neutralized waste may be mixed with sodium hypochlorite in volumes such that the final mixture contains 1.0% sodium hypochlorite. Allow 30 minutes for effective decontamination. Liquid waste containing acid must be neutralized prior to the addition of sodium hypochlorite.

9. PREPARATION OF REAGENTS

- Buffer Concentrates should be brought to room temperature and should be diluted before starting the test procedure.
- If crystals have formed in the Buffer Concentrates, warm them gently until they have completely dissolved.

9.1. Wash Buffer (1x)

Pour entire contents (50 ml) of the Wash Buffer Concentrate (20x) into a clean 1000 ml graduated cylinder. Bring to final volume of 1000 ml with glass-distilled or deionized water. Mix gently to avoid foaming.

Transfer to a clean wash bottle and store at 2° to 25°C. Please note that Wash Buffer (1x) is stable for 30 days.

Wash Buffer (1x) may also be prepared as needed according to the following table:

Number of Strips	Wash Buffer Concentrate (20x) (ml)	Distilled Water (ml)
1 - 6	25	475
1 - 12	50	950

9.2. Assay Buffer (1x)

Pour the entire contents (5 ml) of the Assay Buffer Concentrate (20x) into a clean 100 ml graduated cylinder. Bring to final volume of 100 ml with distilled water. Mix gently to avoid foaming.

Store at 2° to 8°C. Please note that the Assay Buffer (1x) is stable for 30 days.

Assay Buffer (1x) may also be prepared as needed according to the following table:

Number of Strips	Assay Buffer Concentrate (20x) (ml)	Distilled Water (ml)
1 - 6	2.5	47.5
1 - 12	5.0	95.0

9.3. Biotin-Conjugate

Note: The Biotin-Conjugate should be used within 30 minutes after dilution.

Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with Assay Buffer (1x) in a clean plastic tube as needed according to the following table:

Number of Strips	Biotin-Conjugate (ml)	Assay Buffer (1x) (ml)
1 - 6	0.03	2.97
1 - 12	0.06	5.94

9.4. Streptavidin-HRP

Note: The Streptavidin-HRP should be used within 30 minutes after dilution.

Make a 1:200 dilution of the concentrated Streptavidin-HRP solution with Assay Buffer (1x) in a clean plastic tube as needed according to the following table:

Number of Strips	Streptavidin-HRP (ml)	Assay Buffer (1x) (ml)
1 - 6	0.03	5.97
1 - 12	0.06	11.94

9.5. Mouse IL-12p70 Standard

Reconstitute mouse IL-12p70 Standard by addition of distilled water. Reconstitution volume is stated in the Quality Control Sheet. Swirl or mix gently to insure complete and homogeneous solubilization (concentration of reconstituted standard = 2 ng/ml).

Allow the standard to reconstitute for 10-30 minutes. Mix well prior to making dilutions.


The standard has to be used immediately after reconstitution and cannot be stored.

9.5.1 External standard dilution

- Label 7 tubes, one for each standard point: S1, S2, S3, S4, S5, S6, S7
- Then prepare 1:2 serial dilutions for the standard curve as follows:
Pipette 225 μ l of Sample Diluent into each tube.
- Pipette 225 μ l of reconstituted standard (concentration of standard = 2 ng/ml) into the first tube, labelled S1, and mix (concentration of standard 1 = 1 ng/ml).
- Pipette 225 μ l of this dilution into the second tube, labelled S2, and mix thoroughly before the next transfer. Repeat serial dilutions 5 more times thus creating the points of the standard curve (see Figure 6).

Sample Diluent serves as blank.

Figure 6 Dilute standards - tubes



10. TEST PROTOCOL

1. Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running blanks and standards. Each sample, standard, blank and optional control sample should be assayed in duplicate. Remove extra microwell strips from holder and store in foil bag with the desiccant provided at 2°C to 8°C sealed tightly.
2. Wash the microwell strips twice with approximately 400 µL Wash Buffer per well with thorough aspiration of microwell contents between washes. Allow the Wash Buffer to sit in the wells for about 10 – 15 seconds before aspiration. Take care not to scratch the surface of the microwells.

After the last wash step, empty wells and tap microwell strips on absorbent pad or paper towel to remove excess Wash Buffer. Use the microwell strips immediately after washing. Alternatively microwell strips can be placed upside down on a wet absorbent paper for not longer than 15 minutes. Do not allow wells to dry.

3. Standard dilution on the microwell plate (Alternatively the standard dilution can be prepared in tubes - see 9.5.1):

Add 100 µL of Sample Diluent in duplicate to all standard wells. Pipette 100 µL of prepared standard (see “Mouse IL-12p70 standard“, concentration = 2000 pg/mL) in duplicate into well A1 and A2 (see Table 1). Mix the contents of wells A1 and A2 by repeated aspiration and ejection (concentration of standard 1, S1 = 1000 pg/mL), and transfer 100 µL to wells B1 and B2, respectively (see Figure 7). Take care not to scratch the inner surface of the microwells. Continue this procedure 5 times, creating two rows of mouse IL-12p70 standard dilutions ranging from 1000.0 to 15.6 pg/mL. Discard 100 µL of the contents from the last microwells (G1, G2) used.

Figure 7




Table 1

Example of the arrangement of blanks, standards and samples in the microwell strips:

	1	2	3	4
A	Standard 1 1000.0 pg/mL	Standard 1 1000.0 pg/mL	Sample 1	Sample 1
B	Standard 2 500.0 pg/mL	Standard 2 500.0 pg/mL	Sample 2	Sample 2
C	Standard 3 250.0 pg/mL	Standard 3 250.0 pg/mL	Sample 3	Sample 3
D	Standard 4 125.0 pg/mL	Standard 4 125.0 pg/mL	Sample 4	Sample 4
E	Standard 5 62.5 pg/mL	Standard 5 62.5 pg/mL	Sample 5	Sample 5
F	Standard 6 31.3 pg/mL	Standard 6 31.3 pg/mL	Sample 6	Sample 6
G	Standard 7 15.6 pg/mL	Standard 7 15.6 pg/mL	Sample 7	Sample 7
H	Blank	Blank	Sample 8	Sample 8

In case of an external standard dilution (see “External standard dilution”), pipette 100 µL of these standard dilutions (S1 - S7) in the standard wells according to Table 1.

4. Add 100 µL of Sample Diluent in duplicate to the blank wells.
5. Add 50 µL of Sample Diluent to the sample wells.
6. Add 50 µL of each sample in duplicate to the sample wells.
7. Prepare Biotin-Conjugate (see Preparation of Biotin-Conjugate 9.3).
8. Add 50 µL of Biotin-Conjugate to all wells.
9. Cover with an adhesive film and incubate at room temperature (18 to 25°C) for 2 hours on a microplate shaker.
10. Prepare Streptavidin-HRP (refer to Preparation of Streptavidin-HRP 9.4).
11. Remove adhesive film and empty wells. Wash microwell strips 3 times according to point 2 of the test protocol. Proceed immediately to the next step.
12. Add 100 µL of diluted Streptavidin-HRP to all wells, including the blank wells.
13. Cover with an adhesive film and incubate at room temperature (18 to 25°C) for 1 hour on a microplate shaker.
14. Remove adhesive film and empty wells. Wash microwell strips 3 times according to point 2 of the test protocol. Proceed immediately to the next step.
15. Pipette 100 µL of TMB Substrate Solution to all wells.
16. Incubate the microwell strips at room temperature (18 to 25°C) for about 10 min. Avoid direct exposure to intense light.

The color development on the plate should be monitored and the substrate reaction stopped (see next point of this protocol) before positive wells are no longer properly recordable. Determination of the ideal time period for color development has to be done individually for each assay.

It is recommended to add the stop solution when the highest standard has developed a dark blue color. Alternatively the color development can be monitored by the ELISA reader at 620 nm. The substrate reaction should be stopped as soon as Standard 1 has reached an OD of 0.9 – 0.95.

17. Stop the enzyme reaction by quickly pipetting 100 μ l of Stop Solution into each well. It is important that the Stop Solution is spread quickly and uniformly throughout the microwells to completely inactivate the enzyme. Results must be read immediately after the Stop Solution is added or within one hour if the microwell strips are stored at 2 - 8°C in the dark.
18. Read absorbance of each microwell on a spectro-photometer using 450 nm as the primary wave length (optionally 620 nm as the reference wave length; 610 nm to 650 nm is acceptable). Blank the plate reader according to the manufacturer's instructions by using the blank wells. Determine the absorbance of both the samples and the standards.

Note: In case of incubation without shaking the obtained O.D. values may be lower than indicated below. Nevertheless the results are still valid.

11. CALCULATION OF RESULTS

- Calculate the average absorbance values for each set of duplicate standards and samples. Duplicates should be within 20 percent of the mean value.
- Create a standard curve by plotting the mean absorbance for each standard concentration on the ordinate against the mouse IL-12p70 concentration on the abscissa. Draw a best fit curve through the points of the graph (a 5-parameter curve fit is recommended).
- To determine the concentration of circulating mouse IL-12p70 for each sample, first find the mean absorbance value on the ordinate and extend a horizontal line to the standard curve. At the point of intersection, extend a vertical line to the abscissa and read the corresponding mouse IL-12p70 concentration.
- If instructions in this protocol have been followed, samples have been diluted 1:2 (50 μ L sample + 50 μ L Sample Diluent) and the concentration read from the standard curve must be multiplied by the dilution factor (x 2).
- Calculation of 1:2 prediluted samples with a concentration exceeding standard 1 may result in incorrect, low mouse IL-12p70 levels. Such samples require further external predilution according to expected mouse IL-12p70 values with Sample Diluent in order to precisely quantitate the actual mouse IL-12p70 level.
- It is suggested that each testing facility establishes a control sample of known mouse IL-12p70 concentration and runs this additional control with each assay. If the values obtained are not within the expected range of the control, the assay results may be invalid.
- A representative standard curve is shown in Figure 8. This curve cannot be used to derive test results. Each laboratory must prepare a standard curve for each group of microwell strips assayed.

Figure 8

Representative standard curve for mouse IL-12p70 ELISA. Mouse IL-12p70 was diluted in serial 2-fold steps in Sample Diluent. Do not use this standard curve to derive test results. A standard curve must be run for each group of microwell strips assayed.




Table 2
 Typical data using the mouse IL-12p70 ELISA
 Measuring wavelength: 450 nm
 Reference wavelength: 620 nm

Standard	Mouse IL-12p70 Concentration (pg/mL)	O.D. at 450 nm	Mean O.D. at 450 nm	C.V. (%)
1	1000	2.009 1.979	1.994	1.1
2	500	1.166 1.182	1.174	1.0
3	250	0.641 0.657	0.649	1.7
4	125	0.345 0.344	0.345	0.2
5	62.5	0.191 0.212	0.202	7.4
6	31.3	0.116 0.125	0.121	5.3
7	15.6	0.077 0.078	0.078	0.9
Blank	0.0	0.044 0.040	0.042	6.7

The OD values of the standard curve may vary according to the conditions of assay performance (e.g. operator, pipetting technique, washing technique or temperature effects). Furthermore shelf life of the kit may affect enzymatic activity and thus color intensity. Values measured are still valid.

12. LIMITATIONS

- Since exact conditions may vary from assay to assay, a standard curve must be established for every run.
- Bacterial or fungal contamination of either screen samples or reagents or cross-contamination between reagents may cause erroneous results.
- Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.
- Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Empty wells completely before dispensing fresh wash solution, fill with Wash Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

13. PERFORMANCE CHARACTERISTICS

13.1. Sensitivity

The limit of detection of mouse IL-12p70 defined as the analyte concentration resulting in an absorbance significantly higher than that of the dilution medium (mean plus 2 standard deviations) was determined to be < 10 pg/mL (mean of 6 independent assays).

13.2. Reproducibility

13.2.1. Intra-assay

Reproducibility within the assay was evaluated in 3 independent experiments. Each assay was carried out with 6 replicates of 7 serum samples containing different concentrations of mouse IL-12p70. Two standard curves were run on each plate. Data below show the mean mouse IL-12p70 concentration and the coefficient of variation for each sample (see Table 3). The calculated overall intra-assay coefficient of variation was 8.3%.

Table 3

The mean mouse IL-12p70 concentration and the coefficient of variation for each sample

Sample	Experiment	Mean Mouse IL-12p70 Concentration (pg/mL)	Coefficient of variation (%)
1	1	630	9
	2	495	10
	3	600	6
2	1	365	9
	2	352	5
	3	362	3
3	1	475	11
	2	475	5
	3	397	3
4	1	850	8
	2	670	12
	3	858	10
5	1	412	11
	2	330	7
	3	376	8
6	1	219	9
	2	165	7
	3	173	6
7	1	135	14
	2	105	10
	3	126	10

13.2.2. Inter-assay

Assay to assay reproducibility within one laboratory was evaluated in 3 independent experiments. Each assay was carried out with 6 replicates of 7 serum containing different concentrations of mouse IL-12p70. Two standard curves were run on each plate. Data below show the mean mouse IL-12p70 concentration and the coefficient of variation calculated on 18 determination of each samples (see Table 4). The calculated overall inter-assay coefficient of variation was 11%.

Table 4

The mean mouse IL-12p70 concentration and the coefficient of variation of each sample

Sample	Mean Mouse IL-12p70 Concentration (pg/mL)	Coefficient of Variation (%)
1	575	12.3
2	360	1.9
3	449	10.0
4	793	13.4
5	374	11.0
6	186	15.6
7	122	12.5

13.3. Spiking Recovery

The spike recovery was evaluated by spiking 3 levels of mouse IL-12p70 into pooled murine serum. Recoveries were determined in 3 independent experiments with 2 replicates each.

The unspiked serum was used as blank in these experiments.

The recovery ranged from 95% to 118% with an overall mean recovery of 106%

13.4. Dilution Linearity

Serum samples with different levels of mouse IL-12p70 were analyzed at serial 2-fold dilutions with 4 replicates each.

The recovery ranged from 87.7% to 117.1% with an overall recovery of 101.9%.

Table 5

Sample	Dilution	Expected Mouse IL-12p70 Concentration (pg/mL)	Observed Mouse IL-12p70 Concentration (pg/mL)	Recovery of Expected Mouse IL-12p70 Concentration (%)
1	1:2	-	711	-
	1:4	356	385	108.3
	1:8	193	207	107.5
	1:16	104	117	113.1
2	1:2	-	1013	-
	1:4	507	494	97.5
	1:8	247	233	94.1
	1:16	116	118	101.4
3	1:2	-	467	-
	1:4	234	208	89.0
	1:8	104	108	103.8
	1:16	54	56	103.8
4	1:2	-	545	-
	1:4	272	239	87.7
	1:8	119	119	99.3
	1:16	59	69	117.1

13.5. Sample Stability

13.5.1. Freeze-Thaw Stability

Aliquots of serum samples (spiked or unspiked) were stored at -20°C and thawed 5 times, and the mouse IL-12p70 levels determined. There was no significant loss of mouse IL-12p70 immunoreactivity detected by freezing and thawing.

13.5.2. Storage stability

Aliquots of serum samples (spiked or unspiked) were stored at -20°C, 2°C to 8°C, room temperature, and at 37°C, and the mouse IL-12p70 level determined after 24 hours. There was no significant loss of mouse IL-12p70 immunoreactivity detected during storage under above conditions.

13.6. Specificity

The cross reactivity and interference of circulating factors of the immune system was evaluated by spiking these proteins at physiologically relevant concentrations into a mouse IL-12p70 positive sample.

There was no cross reactivity or interference detected.

13.7. Expected Values

Serum samples from randomly selected apparently healthy mice were tested for mouse IL-12p70. There were no detectable mouse IL-12p70 levels found. Elevated mouse IL-12p70 levels depend on the type of immunological disorder.

14. REAGENT PREPARATION SUMMARY

14.1 Wash Buffer (1x)

Add Wash Buffer Concentrate 20x (50 ml) to 950 ml distilled water.

Number of Strips	Wash Buffer Concentrate (ml)	Distilled Water (ml)
1 – 6	25	475
1 – 12	50	950

14.2 Assay Buffer (1x)

Add Assay Buffer Concentrate 20x (5 ml) to 95 ml distilled water.

Number of Strips	Assay Buffer Concentrate (ml)	Distilled Water (ml)
1 – 6	2.5	47.5
1 – 12	5.0	95.0

14.3 Biotin-Conjugate

Make a 1:100 dilution of Biotin-Conjugate in Assay Buffer (1x):

Number of Strips	Biotin-Conjugate (ml)	Assay Buffer (1x) (ml)
1 – 6	0.03	2.97
1 – 12	0.06	5.94

14.4 Streptavidin-HRP

Make a 1:200 dilution of Streptavidin-HRP in Assay Buffer (1x):

Number of Strips	Streptavidin-HRP (ml)	Assay Buffer (1x) (ml)
1 - 6	0.03	5.97
1 - 12	0.06	11.94

14.5 Mouse IL-12 p70 Standard


Reconstitute lyophilized mouse IL-12p70 standard with distilled water. Reconstitution volume is stated in the Quality Control Sheet.


15. TEST PROTOCOL SUMMARY

1. Determine the number of microwell strips required.
2. Wash microwell strips twice with Wash Buffer.
3. Standard dilution on the microwell plate: Add 100 µl Sample Diluent , in duplicate, to all standard wells. Pipette 100 µl prepared standard into the first wells and create standard dilutions by transferring 100 µl from well to well. Discard 100 µl from the last wells.
Alternatively external standard dilution in tubes (see 9.5.1): Pipette 100 µl of these standard dilutions in the microwell strips.
4. Add 100 µl Sample Diluent, in duplicate, to the blank wells.
5. Add 50 µl Sample Diluent to sample wells.
6. Add 50 µl sample in duplicate, to designated sample wells.
7. Prepare Biotin-Conjugate.
8. Add 50 µl Biotin-Conjugate to all wells.
9. Cover microwell strips and incubate 2 hours at room temperature (18 to 25°C).
10. Prepare Streptavidin-HRP.
11. Empty and wash microwell strips 3 times with Wash Buffer.
12. Add 100 µl diluted Streptavidin-HRP to all wells.
13. Cover microwell strips and incubate 1 hour at room temperature (18 to 25°C).
14. Empty and wash microwell strips 3 times with Wash Buffer.
15. Add 100 µl of TMB Substrate Solution to all wells.
16. Incubate the microwell strips for about 10 minutes at room temperature (18 to 25°C).
17. Add 100 µl Stop Solution to all wells.
18. Blank microwell reader and measure colour intensity at 450 nm.

Note: If instructions in this protocol have been followed, samples have been diluted 1:2 (50 µL sample + 50 µL Sample Diluent) and the concentration read from the standard curve must be multiplied by the dilution factor (x 2).

NOTES





BioVendor – Laboratorní medicína, a.s.

Karasek 1767/1, 621 00 Brno, Czech Republic

Phone: +420-549-124-185, Fax: +420-549-211-460

E-mail: info@biovendor.com, sales@biovendor.com

Web: www.biovendor.com

There are BioVendor branches and distributors near you.

To find the office closest to you, visit www.biovendor.com/contact

